

How to Switch

When we changed trigger to 'coincidence' after an alphas calibration we did not unselect 'secondary' in the S800 Trigger GUI. This was run 229, and the dead time was quite large